



**Australian Government**  
**Department of Health and Ageing**



# Horizon Scanning Technology Prioritising Summary

The IntraLase® femtosecond laser

September 2006  
(Updated October 2007 and November 2008)



**Australian  
Safety  
and Efficacy  
Register  
of New  
Interventional  
Procedures -  
Surgical**



**Royal Australasian  
College of Surgeons**



© Commonwealth of Australia [2006]

This work is copyright. You may download, display, print and reproduce this material in unaltered form only (retaining this notice) for your personal, non-commercial use or use within your organisation. Apart from any use as permitted under the Copyright Act 1968, all other rights are reserved. Requests and inquiries concerning reproduction and rights should be addressed to Commonwealth Copyright Administration, Attorney General's Department, Robert Garran Offices, National Circuit, Canberra ACT 2600 or posted at <http://www.ag.gov.au/cca>

Electronic copies can be obtained from <http://www.horizonscanning.gov.au>

Enquiries about the content of the report should be directed to:

HealthPACT Secretariat  
Department of Health and Ageing  
MDP 106  
GPO Box 9848  
Canberra ACT 2606  
AUSTRALIA

**DISCLAIMER:** This report is based on information available at the time of research and cannot be expected to cover any developments arising from subsequent improvements to health technologies. This report is based on a limited literature search and is not a definitive statement on the safety, effectiveness or cost-effectiveness of the health technology covered.

The Commonwealth does not guarantee the accuracy, currency or completeness of the information in this report. This report is not intended to be used as medical advice and it is not intended to be used to diagnose, treat, cure or prevent any disease, nor should it be used for therapeutic purposes or as a substitute for a health professional's advice. The Commonwealth does not accept any liability for any injury, loss or damage incurred by use of or reliance on the information.

The production of this Horizon scanning prioritising summary was overseen by the Health Policy Advisory Committee on Technology (HealthPACT), a sub-committee of the Medical Services Advisory Committee (MSAC). HealthPACT comprises representatives from health departments in all states and territories, the Australia and New Zealand governments; MSAC and ASERNIP-S. The Australian Health Ministers' Advisory Council (AHMAC) supports HealthPACT through funding.

This Horizon scanning prioritising summary was prepared by staff from the Australian Safety and Efficacy Register of New Interventional Procedures – Surgical (ASERNIP-S).

**Name of Technology:**

IntraLase® femtosecond laser

**Purpose and Target Group:**

The IntraLase femtosecond laser was developed to create corneal flaps required for laser in situ keratomileusis (LASIK).

**Stage of Development (in Australia):**

- Experimental
- Investigational
- Nearly established
- Established
- Established but changed indication or modification of technique
- Should be taken out of use
- Not yet emerged

IntraLase is currently available in Australia and hence is registered in the Australian Register of Therapeutic Goods (ARTG number: 107191 and 124974, Product number: 186270 and 208204).

**International Utilisation:**

| COUNTRY       | LEVEL OF USE    |             |                 |
|---------------|-----------------|-------------|-----------------|
|               | Trials underway | Limited use | Widely diffused |
| United States |                 |             | ✓               |
| Europe        |                 |             | ✓               |

**Impact Summary:*****Background***

Laser in situ keratomileusis (LASIK) has emerged to become the most common refractive surgical procedures for correcting myopia, hyperopia and astigmatism since its introduction in the early 1990s. The LASIK procedure is essentially a two-step process, the first step requires the surgeon to create a flap of corneal tissue which is folded back to prepare the eye for the second step. The second step involves the use of an excimer laser which reshapes the inner cornea to correct the vision of the patient (IntraLase Corp. 2006). Despite the high success rates of LASIK there are some potential risks when this procedure is utilised. Studies have revealed that most major vision-threatening complications (e.g. epithelial defects, incomplete flaps, decentered flaps) are linked to the creation of the corneal flap (Binder 2006). Since the introduction of LASIK, various mechanical microkeratomes have been developed to create the corneal flap. However, despite the improvements which have been



made towards the design of microkeratomes, there are several complications which have not been adequately addressed. Flap dimensions created with existing microkeratome technology continues to produce unpredictable ranges of flap thickness that are unrelated to attempted flap thickness. This complication has the potential to place the laser ablation deeper than intended and therefore increasing the risk of corneal ectasia (progressive corneal thinning and bulging). In addition to this, several studies have revealed that current microkeratome systems create thinner flaps when the preoperative corneal thickness is thin and/or when the preoperative corneal curvature exceeds 46 diopters (D) (Binder 2006).

The femtosecond laser is the first alternative to mechanical microkeratomes and the IntraLase is the first femtosecond laser approved by the Food and Drugs Administration in the United States. The IntraLase system employs the precision of a computer-guided laser to create the corneal flap and therefore claims to be more precise compared to hand-held microkeratomes. It utilises an infrared beam of light to separate tissue via a process known as photodisruption. Photodisruption is the process whereby focused laser pulses divide material at the molecular level without the transfer of heat or impact to the surrounding tissue. Unlike mechanical microkeratomes, IntraLase creates the flap below the surface of the cornea using an 'inside-out' process. Each pulse of the laser creates a tiny bubble (2 to 3 microns) of carbon dioxide and water vapour. The laser essentially moves back and forth across the eye, creating a uniform layer of bubbles beneath the corneal surface. Bubbles are then stacked along the edge of the flap up to the corneal surface, thus completing the flap. This level of accuracy is the main reason for the purported safety of IntraLase compared to microkeratomes (IntraLase Corp. 2006). In addition to corneal flap creation, recent studies (Sikder and Snyder 2006, Terry *et al.* 2005) have examined the possibility of utilising IntraLase as a means of obtaining smooth donor cornea for corneal transplantation. However, corneal transplantation is beyond the scope of the current summary and the low-level evidence (*in vitro* and cadaver studies) currently available hinders its inclusion in the following discussion.

### ***Clinical Need and Burden of Disease***

The 2001 National Health Survey reported that 22% (~ 4 million) of Australians are hyperopic while 5% (~ 1 million) experience astigmatism (AIHW 2004). Most individuals afflicted by hyperopia and astigmatism would be satisfied with corrections using prescription glasses or contact lenses. However, some of these individuals may opt to undergo LASIK surgery to achieve permanent vision correction.

Previous studies have revealed that corneal flap-related complications occur in as many as 5% of LASIK cases and includes decentered flaps, free flaps, irregular flap edges and stromal bed surfaces, epithelial abrasions, buttonholes and flap lacerations. These complications can result in delayed recovery of visual acuity while severe cases may lead to permanent vision loss (Lim *et al.* 2006).



### ***Estimated Speed and Geographic and Practitioner Use Patterns of Diffusion in the Health System***

In Australia, IntraLase has been approved by the TGA and is currently used for corneal flap creation for LASIK procedures. However, the extent of use of this technology in Australia is not documented. The IntraLase laser was approved by the FDA for laser assisted creation of the corneal flap in 2002 and has been used extensively in the United States with approximately 500,000 procedures as of October 2005. In addition to this, IntraLase has received the CE mark of approval in Europe with approximately 30 sites utilising this technology (Centre for sight 2006).

### ***Existing Comparators***

The main comparators to IntraLase are the current mechanical microkeratomes. However it is important to note that Ziemer Ophthalmic Systems AG (Switzerland) had developed the Da Vinci femtosecond surgical laser, but clinical studies are limited at the time of writing (Ziemer Ophthalmics 2006).

### ***Estimated Cost Impact***

LASIK surgery with IntraLase typically costs more than conventional LASIK with a microkeratome. The cost increases by approximately US\$250 to US\$500 per eye, averaging US\$344 (Segre and Thompson 2006). In Australia, the average cost increase of using IntraLase during LASIK surgery is approximately \$400 (Lasik Vision 2006).

### ***Efficacy and Safety Issues***

#### **List of Studies Found**

|                                    |    |
|------------------------------------|----|
| Total number of studies            | 17 |
| Randomised controlled trials       | 1  |
| Non-randomised comparative studies | 2  |
| Case series studies                | 12 |
| Case reports                       | 2  |

The studies included in this summary are highlighted in bold in the reference list. One randomised controlled trial, two non-randomised comparative studies and one case series were included in this summary. The case series by Binder (2006) was selected for inclusion above other case series studies due to the sheer number of corneal flaps created (n = 1000) which offers substantial insight to the efficacy and safety of IntraLase.



## Safety

Achieving accurate corneal flap thickness is a key safety consideration for the prevention of ectasia following LASIK surgery. Compared to mechanical microkeratomes, IntraLase achieved a mean flap thickness 16  $\mu\text{m}$  less than programmed, meanwhile the Carriazo-Barraquer (CB) microkeratome achieved a mean flap thickness that was 23  $\mu\text{m}$  thicker than plate\* thickness and the Hansatome achieved a mean 24  $\mu\text{m}$  thinner than plate thickness. Mean flap thickness was significantly thinner ( $p < 0.001$ ) and the variance of flap thickness was significantly better ( $P < 0.001$ ) with IntraLase compared to the mechanical microkeratomes. In addition to this, no eye in the IntraLase group experienced loose epithelium. In comparison, loose epithelium was evident in 1 quadrant in 5.5% of eyes and 2 quadrants in 4.1% of eyes in the CB group, resulting in an overall rate of 9.6% ( $p = 0.001$ , compared to the IntraLase group). Meanwhile in Hansatome eyes, loose epithelium in 1 quadrant was seen in 5.1% of eyes and loose epithelium in 2 quadrants was reported in 2.6% of eyes for a total of 7.7% ( $p = 0.001$ , compared to the IntraLase group). There were no incidences of buttonholes, transacted flaps or other sight-threatening complications. Similar rates of best spectacle-corrected visual acuity (BSCVA) loss were documented for all three devices (IntraLase, CB, Hansatome) (Kerizian and Stonecipher 2004).

Eyes with corneal flaps created with IntraLase exhibited more post-operative inflammation compared to the mechanical keratome group in the study conducted by Lim *et al.* (2006); however the degree of inflammation was not quantified. In one eye, a mechanical failure caused the inadvertent release of the vacuum during flap creation. This required the suction ring to be remounted and the flap was created successfully on the second pass on the laser (Lim *et al.* 2006).

Binder (2006) documented several operative complications over the course of 1000 consecutive IntraLase corneal flaps. Six eyes experienced a loss of suction that precluded the completion of the operation on the same surgical day. Two incomplete laser passes due to suction loss were reoperated on the same day. Two flaps (in the 90  $\mu\text{m}$  attempted thickness group) tore at the hinge due to excessive tension on the flap caused by the mechanical lift. No flap suffered from epithelial defects, but the peripheral host epithelium of three eyes experienced focal defects at the end of the procedure which was due to trauma at flap lift. No incidences of free, button hole, or decentered flaps were reported even when a previous mechanical microkeratome flap button hole case was reoperated. There were no incidences of visually significant interface inflammation (diffuse lamellar keratitis, DLK) on the first post-operative day, however 50 eyes had minor peripheral interface inflammation (Grade 1 or trace) and were treated with hourly topical prednisone acetate (1%) and cleared in 1 to 3 days. Four eyes had visually significant folds in the Bowman's layer (microstriae), while several eyes (exact numbers not stated) which had no visual symptoms developed a granular,

---

\* A depth plate in microkeratomes determines the planned thickness for the flap resection.



peau d'orange appearance on the interface approximately 3 months post-surgery. Moderate to severe photophobia was documented in 5 patients 4 to 8 weeks following surgery with no loss of best spectacle-corrected visual acuity, no internal or external inflammation was detected in these eyes. Photophobia was resolved with constant topical steroid treatment for 1 to 2 weeks (Binder 2006).

Prior to the software update for IntraLase in 2004, Binder (2006) reported that the mean achieved flap thickness for IntraLase (at the 90  $\mu\text{m}$  setting) exceeded the intended thickness by 15  $\mu\text{m}$  to 30  $\mu\text{m}$ . However, after the software update the mean achieved flap thickness improved substantially and was within 3  $\mu\text{m}$  to 5  $\mu\text{m}$  of the attempted flap thickness. Binder (2006) claims that the standard deviation of flap thickness remained consistent for almost all the intended flap thicknesses in IntraLase eyes, and the range of flap thickness was smaller compared to data reported with mechanical microkeratomes. However, no data was presented to support this statement. Mechanical microkeratomes has the tendency to create thicker flaps when operating on thicker corneas. In comparison, IntraLase results from this study revealed no difference in the achieved flap thickness (for attempted flap thickness of 90  $\mu\text{m}$  and 100  $\mu\text{m}$ ) when comparing preoperative central corneal thickness of less than 500  $\mu\text{m}$  with those greater than 576  $\mu\text{m}$ . Another disadvantage of mechanical microkeratomes is the increased risk of creating thinner flaps when preoperative corneal curvature is greater than 46 D, Binder (2006) reported no difference in achieved flap thickness with the IntraLase despite preoperative mean corneal curvatures of less than 42 D or greater than 46 D (Binder 2006). Meanwhile the study by Tran *et al.* (2005) did not report of any safety issues throughout the study period. These results therefore indicate that IntraLase appears to be capable of producing consistent flap thickness regardless of pre-operative corneal thickness or curvature.

### ***Efficacy***

The randomised controlled trial by Tran *et al.* (2005) attempted to identify the potential advantages of IntraLase compared to a mechanical microkeratome, Hansatome, in wavefront-guided LASIK procedures. One of the key differences in wavefront-guided LASIK ablations compared to conventional LASIK is that the primary data input for ablation is objective data obtained from an aberrometer while conventional ablation utilises subjective refraction data. Due to the fact that aberrometer measurements are done prior to flap creation, any refractive changes that occur after flap creation cannot be programmed into the excimer laser. Therefore it would be advantageous to ensure that flap creation does not introduce any changes that may affect the accuracy of wavefront-guided LASIK. Nine patients with myopia or myopic astigmatism underwent a 2-step procedure with excimer laser treatment 10 weeks after initial flap creation. One eye in each patient was randomised to Hansatome keratectomy and the other eye to IntraLase. Tran *et al.* (2005) showed that 10 weeks after flap creation, statistically significant changes in defocus wavefront aberrations



were observed after IntraLase (baseline: 3.5 RMS, 10 weeks post-flap: ~3.2 RMS,  $p = 0.008$ ) and Hansatome (baseline: 3.5 RMS, 10 weeks post-flap: ~2.7 RMS,  $p = 0.004$ ) flap creation. A hyperopic shift in manifest refraction was seen in the Hansatome group after the creation of the corneal flap ( $p = 0.04$ ), however no statistically significant change in manifest refraction was observed in the IntraLase group. In the Hansatome group, statistically significant increases of higher-order aberrations (trefoil and quadrafoil Zernicke terms) were noted after flap creation ( $p = 0.02$ ). Conversely, no significant changes in higher-order aberrations were observed in the IntraLase group. Differential wavefront maps indicated that eyes with Hansatome-created flaps had significant aberrations while nearly flat wavefront changes were observed in the fellow eyes of the same patients which had IntraLase-created flaps (Tran *et al.* 2005).

A retrospective case study evaluated one surgeon's use of mechanical microkeratomes in 2002, using Carriazo-Barraquer (CB) ( $n = 126$  eyes in 126 patients) or Hansatome ( $n = 146$  eyes in 143 patients), compared with his use of IntraLase ( $n = 106$  eyes in 106 patients) in 3 months in 2003. (Kerizian and Stonecipher 2004). 20/20 uncorrected visual acuity (UCVA) rates at day 1 post-surgery was slightly lower in the IntraLase group (45%) compared to CB (57%) and Hansatome (51%), but this was not statistically significant ( $p$ -value not stated). Similarly, when the 20/40 UCVA rate was considered, all groups achieved comparable results (IntraLase: 98%, CB: 99%, Hansatome: 99%) (Kerizian *et al.* 2004). Mean acuity for all groups was similar at day 1. Three months after surgery, 20/20 UCVA rates remained comparable in all three groups; 67% for IntraLase, 71% for CB and 66% for Hansatome. Meanwhile 20/40 rates 3 months post-surgery ranged from 98% to 99% (IntraLase: 98%, CB: 99%, Hansatome: 99%). Refractive predictability was significantly better for the IntraLase eyes, with 91% (95% CI, 88% to 94%) achieving a manifest refractive spheroequivalent (MRSE) of  $\pm 0.50$  D, while the MRSE  $\pm 0.50$  D rates for CB and Hansatome was 73% (95% CI, 69.4% to 77.4%;  $p < 0.01$ ) and 74% (95% CI, 69.9% to 77.3%;  $p < 0.01$ ) respectively. Meanwhile, all three groups achieved similar rate of MRSE  $\pm 1.00$  D; IntraLase 99%, CB 99% and Hansatome 95%. Linear regression analysis of the attempted MRSE versus the achieved MRSE revealed that IntraLase eyes had significantly better correlation compared to the CB group ( $p < 0.05$ ) but not for the Hansatome group ( $p = 0.14$ ). Surgically induced astigmatism in spherical corrections is a common occurrence in LASIK surgery. In this study, Kerizian and Stonecipher (2004) reported that IntraLase flaps induced significantly less astigmatism compared to flaps created with CB and Hansatome ( $p < 0.01$ ). The standard deviation of postoperative cylinder in eyes that had spherical corrections (i.e. no cylinder treated) was 0.22 D for IntraLase, 0.32 D for CB and 0.40 D for Hansatome eyes ( $p < 0.01$  for IntraLase vs CB and Hansatome). The authors proposed that the lack of procedure-induced astigmatism is due to the circular (rather than truncated) shape of the IntraLase flap. Other potential contributors include the programmed edge angle and constant flap thickness; however further investigation is necessary (Kerizian and Stonecipher 2004).



Lim *et al.* (2006) compared LASIK outcomes when the corneal flap was created with IntraLase (n = 28 eyes in 14 patients) or Hansatome (n = 27 eyes in 16 patients). Patients were allocated to either procedure according to their choice. Pre-operative spherical error for IntraLase eyes was  $-5.6 \pm 2.3$  D (range -2.5 to -10.75 D) while Hansatome eyes was  $-4.9 \pm 1.6$  D (range from -2.5 to -7.75 D). Meanwhile, pre-operative cylindrical error was  $0.8 \pm 0.9$  D (range 0 to 2.5 D) in the IntraLase group and  $0.6 \pm 0.5$  D (range 0 to 1.35 D) in the Hansatome group. Overall, pre-operative refractive errors were similar between both groups ( $p > 0.05$ ). One month post-surgery, mean spherical error was comparable for both groups, with mean  $\pm$  SD of  $-0.40 \pm 0.5$  D for IntraLase eyes and  $-0.40 \pm 0.6$  D for Hansatome eyes. In the same way, post-operative mean cylindrical error at one month post-surgery was comparable between both groups as well, with a mean of  $0.20 \pm 0.4$  D for the IntraLase group and  $0.30 \pm 0.5$  D for the Hansatome group. The similarity between IntraLase and Hansatome eyes for spherical and cylindrical errors was maintained up to 3 months post-surgery (Table 2), indicating that the level of procedure-induced astigmatism was similar between both groups. These results were contradictory to Kerizian and Stonecipher (2004) where procedure-induced astigmatism was significantly lower for IntraLase eyes compared to eyes with flaps created by microkeratomers.

**Table 2: Post-operative spherical and cylindrical errors for IntraLase and Hansatome groups (Lim *et al.* 2006).**

| Error                          | Hansatome group (n = 27) |            | IntraLase group (n = 28) |            |
|--------------------------------|--------------------------|------------|--------------------------|------------|
|                                | 1 month                  | 3 months   | 1 month                  | 3 months   |
| Spherical error < 0.5 D        | 19 (70.4%)               | 20 (74.1%) | 17 (60.7%)               | 18 (64.3%) |
| Spherical error $\geq$ 0.5 D   | 8 (29.6%)                | 7 (25.9%)  | 11 (39.3%)               | 10 (35.6%) |
| Cylindrical error < 0.5 D      | 17 (63.0%)               | 19 (70.4%) | 23 (82.1%)               | 21 (75.0%) |
| Cylindrical error $\geq$ 0.5 D | 10 (37%)                 | 8 (29.6%)  | 5 (17.9%)                | 7 (25.0%)  |

\* Expressed as number of eyes (%)

Graphical representation of the relationship between achieved correction versus intended correction revealed that IntraLase showed a tendency towards undercorrection in myopia < 6 D. Lim *et al.* (2006) reported no loss of BCVA in either group after surgery. In addition, no statistically significant difference was noted in term of postoperative visual acuity between the two groups ( $p > 0.05$ ) (Table 3).



**Table 3: Postoperative uncorrected visual acuity for IntraLase and Hansatome groups (Lim *et al.* 2006)**

| Uncorrected visual acuity | >20/15    | 20/20      | <20/25    |
|---------------------------|-----------|------------|-----------|
| <b>Intralase (n = 28)</b> |           |            |           |
| 1 month                   | 2 (7.1%)  | 21 (75%)   | 5 (17.9%) |
| 3 months                  | 4 (14.3%) | 18 (64.3%) | 6 (21.4%) |
| <b>Hansatome (n = 27)</b> |           |            |           |
| 1 month                   | 1 (3.7%)  | 19 (70.4%) | 7 (25.9%) |
| 3 months                  | 2 (7.4%)  | 20 (74.2%) | 5 (18.5%) |

\* Expressed as number of eyes (%)

It is a generally accepted fact that after LASIK surgery, higher order aberrations increase due to the procedure. In this study, root mean square of higher-order aberrations (HO-RMS) increased from a preoperative value of  $0.314 \pm 0.230 \mu\text{m}$  to  $0.371 \pm 0.225 \mu\text{m}$  3 months post-surgery in the IntraLase group, while the Hansatome eyes documented an increase from  $0.305 \pm 0.171 \mu\text{m}$  to  $0.430 \pm 0.222 \mu\text{m}$  3 months post-surgery, therefore a significant increase of HO-RMS was noted in both groups ( $p < 0.05$  for both groups). Previous studies have inferred that this event was due to the use of microkeratomes; however in the current comparative study by Lim *et al.* (2006), the increase of higher order aberrations were similar between both groups ( $p > 0.05$ ), an outcome which is contradictory to the randomised control trial by Tran *et al.* (2005). However there was one exception, significantly higher spherical aberrations were reported in the Hansatome group compared to the IntraLase group ( $p < 0.05$ ), an attribute which the authors believe was due to the flap configuration (square-shaped edge) of IntraLase. Post-operative contrast sensitivity was similar for both groups under photopic conditions. Yet under mesopic conditions, contrast sensitivity increased in both groups, with significant improvements in IntraLase eyes at high spatial frequencies (12 and 18 cycles per degree) ( $p \leq 0.05$ ). This result was unexpected as previous studies reported decreases in contrast sensitivity after LASIK. With regards to corneal sensitivity, the evaluation of 10 eyes from each group revealed decreased sensitivity post-surgery. Both groups exhibited significantly slower improvement of corneal sensitivity at the central area compared to the peripheral areas ( $p \leq 0.05$ ). However, the recovery of corneal sensitivity was significantly faster in IntraLase eyes for the central and inferior corneal areas ( $p < 0.05$ ) (Lim *et al.* 2006).

### ***Ethical Issues***

No issues were identified from the retrieved material.

### ***Cultural or Religious Considerations***

No issues were identified from the retrieved material.



## **Other Issues**

Preliminary investigations on the utility of the IntraLase laser in corneal transplantation reveals that it is capable of generating thin donor buttons for transplantation of corneal endothelium. However, Sikder and Snyder (2006) reported that IntraLase caused collateral damage to endothelial cells, indicating that further research is required to optimise endothelial cell survival before IntraLase can be used successfully in corneal transplantation procedures. In an earlier cadaver study, Terry *et al.* (2005) reported that the histology of the IntraLase-formed stromal dissections did not appear substantially better compared to manual deep lamellar endothelial keratoplasty dissections in both recipient and donor tissues. Therefore, additional research is required to examine and fine-tune the potential use of IntraLase in corneal transplantation surgery.

## **Recommendation**

Current studies demonstrate that IntraLase is capable of producing predictable flap thickness, increased post-operative flap stability and decreased epithelial injury (Binder 2006, Kerizian and Stonecipher 2004). However, there are contentious results with regards to the benefits of IntraLase towards improving postoperative refractive errors (Kerizian and Stonecipher 2004, Lim *et al.* 2006), while improvements in clinical outcomes such as visual acuities and total higher-order aberrations were not apparent in the studies discussed. The potential of utilising IntraLase in corneal transplantation is of significant interest, however the issues identified in the current early-stage studies reveals that further research is required before IntraLase can be utilised in clinical corneal transplantation. Based on the evidence available, it is recommended that IntraLase is monitored for further developments in corneal transplantation.

- |  |  |
|--|--|
| <input type="checkbox"/> Horizon Scanning Report | <input type="checkbox"/> Full Health Technology Assessment |
| <input checked="" type="checkbox"/> Monitor      | <input type="checkbox"/> Archive                           |

## **References:**

AIHW – Australia’s Health 2004. Last updated 2004.

<http://www.aihw.gov.au/publications/index.cfm/title/10014> [Accessed July 2006].

**Binder PS. One thousand consecutive IntraLase laser in situ keratomileusis flaps. *Journal of Cataract and Refractive Surgery* 2006; 32(6): 962-969.**

Centre for Sight. IntraLase femtosecond laser track record. Last updated 2006.

<http://www.centreforsight.com/trackrecord.php> [Accessed July 2006].

Intralase Corporation. Last updated 2006. <http://www.intralase.com/> [Accessed July 2006].



**Kerizian GM, Stonecipher KG. Comparison of the IntraLase femtosecond laser and mechanical keratomes for laser in situ keratomileusis. *Journal of Cataract and Refractive Surgery* 2004; 30(4): 804-811.**

Lasik Vision. Cost of LASIK eye surgery. Last updated 2006.  
<http://www.lasikvision.com.au/cost-lasik-surgery.html> [Accessed July 2006].

**Lim T, Yang S, Kim M, Tchah H. Comparison of the IntraLase femtosecond laser and mechanical microkeratome for laser in situ keratomileusis. *The American Journal of Ophthalmology* 2006; 141(5): 833-839.**

Medicare Australia. Medicare Benefits Schedule. Last updated 2006.  
<http://www9.health.gov.au/mbs/> [Accessed July 2006].

Segre L and Thompson V. Cost of LASIK and other corrective surgery. Last updated 2006.  
<http://www.allaboutvision.com/visionsurgery/cost.htm> [Accessed July 2006].

Terry MA, Ousley PJ, Will B. A practical femtosecond laser procedure for DLEK endothelial transplantation. *Cornea* 2005; 24(4): 453-459.

**Tran DB, Sarayba MA, Bor Z, Garufis C, Dyj YJ, Soltes CR, Juhasz T, Kurtz RM. Randomized prospective clinical study comparing induced aberrations with IntraLase and Hansatome flap creation in fellow eyes: potential impact on wavefront-guided laser in situ keratomileusis. *Journal of Cataract and Refractive Surgery* 2005; 31(1): 97-105.**

Ziemer Ophthalmics. Last updated 2006. <http://www.ziemer-ophthalmics.com/> [Accessed August 2006]

#### **Sources of further information:**

Chung SH, Roh SI, Park MS, Kong YT, Lee HK, Kim EK. Mycobacterium abscessus keratitis after LASIK with IntraLase femtosecond laser. *Ophthalmologica* 2006; 220(4): 277-280.

Durrie DS, Kerizian GM. Femtosecond laser versus mechanical keratome flaps in wavefront-guided laser in situ keratomileusis: prospective contralateral eye study. *Journal of Cataract and Refractive Surgery* 2005; 31(1): 120-126.

Eisner RA, Binder PS. Technique for measuring laser in situ keratomileusis flap thickness using the IntraLase laser. *Journal of Cataract and Refractive Surgery* 2006; 32(4): 556-568.

Mahler O, Hirsh A, Kremer I, Barequet IS, Marcovich AL, Nemet P, Levinger S. Laser in situ keratomileusis in myopic patients with congenital nystagmus. *Journal of Cataract and Refractive Surgery* 2006; 32(3): 464-467.



- Munoz G, Albarran-Diego C, Sakla HF, Perez-Santoja JJ, Alio JL. Femtosecond laser in situ keratomileusis after radial keratotomy. *Journal of Cataract and Refractive Surgery* 2006; 32(8): 1270-1275.
- Petroll WM, Goldberg D, Lindsey SS, Kelley PS, Cavanagh HD, Bowman RW, Parmar DN, Verity SM, McCulley JP. Confocal assessment of the corneal response to intracorneal lens insertion and laser in situ keratomileusis with flap creation using Intralase. *Journal of Cataract and Refractive Surgery* 2006; 32(7): 1119-1128.
- Principe AH, Lin DY, Small KW, Aldave AJ. Macular hemorrhage after in situ keratomileusis (LASIK) with femtosecond laser flap creation. *The American Journal of Ophthalmology* 2004; 138(4): 657-659.
- Ratkay-Traub I, Ferincz IE, Juhasz T, Kurtz RM, Krueger RR. First clinical results with the femtosecond neodymium-glass laser in refractive surgery. *Journal of Refractive Surgery* 2003; 19(2): 94-103.
- Sikder S, Snyder RW. Femtosecond laser preparation of donor tissue from the endothelial side. *Cornea* 2006; 25(4): 416-422.
- Sonigo B, Chong Sit D, Ancel JM, Auclin F, Bokobza Y, Baudouin C. In vivo confocal microscopy evaluation of corneal changes induced after LASIK using the IntraLase femtosecond laser. *Journal Francais d'Ophthalmologie* 2005; 28(5): 463-472.
- Stonecipher KG, Dishler JG, Ignacio TS, Binder PS. Transient light sensitivity after femtosecond laser flap creation: clinical findings and management. *Journal of Cataract and Refractive Surgery* 2006; 32(1): 91-94.
- Stonecipher K, Ignacio TS, Stonecipher M. Advances in refractive surgery: microkeratome and femtosecond laser flap creation in relation to safety, efficacy, predictability, and biomechanical stability. *Current Opinion in Ophthalmology* 2006; 17(4): 368-372.
- Talamo JH, Meltzer J, Gardner J. Reproducibility of flap thickness with Intralase FS and Moria LSK-1 and M2 microkeratomes. *Journal of Refractive Surgery* 2006; 22(6): 556-561.
- Touboul D, Salin F, Mortemousque B, Chabassier P, Mottay E, Leger F, Colin J. Advantages and disadvantages of the femtosecond laser microkeratome. *Journal Francais d'Ophthalmologie* 2005; 28(5): 535-546.

**Search Criteria:**

A search of MEDLINE, PubMed, *The Cochrane Library*, the Current Controlled Trials metaRegister, the UK National Research Register, the International Network of Agencies for Health Technology Assessment, relevant online journals and the Internet was conducted in July 2006.

Search terms used were: 'IntraLase', 'femtosecond laser', 'laser cornea\$ flap', 'laser LASIK', 'cornea\$ flap creation'.

---

This Horizon Scanning Prioritising Summary was prepared by Mr. Irving Lee from the NET-S Project, ASERNIP-S for the Health Policy Advisory Committee on Technology (Health PACT), on behalf of the Medical Services Advisory Committee (MSAC) and the Australian Health Ministers' Advisory Council (AHMAC).



## 2007 update

A search of relevant databases, online journals and the Internet was conducted in October 2007, following the recommendation in September 2006 that IntraLase be monitored for 12 months particularly for any developments on the use of this technology in corneal transplantation. A total of seven new studies on the utilisation of IntraLase for LASIK and corneal transplantation were retrieved and included in this update.

### Safety and effectiveness issues

#### a) Safety

None of the new retrieved studies reported incidences of procedural complications or adverse events relating to the use of IntraLase.

#### b) Efficacy

#### *LASIK*

The randomised controlled trial by Patel et al. (2007) compared corneal haze and visual outcomes between fellow eyes that were randomised to LASIK with IntraLase flap creation (bladeless LASIK) or mechanical microkeratomes flap creation. Corneal backscatter from the middle third of the cornea was 6% higher after bladeless LASIK at 1 month post-treatment compared to LASIK with the mechanical microkeratome ( $p = 0.007$ ). No differences were noted at 3- and 6-months post-treatment. Contrast sensitivity, high-contrast visual acuity, refractive error and intraocular forward light scatter did not differ significantly between bladeless LASIK and microkeratome LASIK at any point of examination. At 1-month post-treatment, bladeless LASIK flap thickness was  $143 \pm 16\mu\text{m}$ , thicker than the intended  $120\mu\text{m}$  ( $p < 0.001$ ); while microkeratome LASIK flap thickness was  $138 \pm 22\mu\text{m}$ , thinner than the intended  $180\mu\text{m}$  ( $p < 0.001$ ). There were no statistical difference in variances between bladeless LASIK and microkeratome LASIK. Overall, the results indicate that the method of flap creation did not affect visual outcomes at least up to 6 months post-LASIK treatment. Corneal backscatter was significantly higher for bladeless LASIK in the early stages of follow-up, however patients did not notice any difference in vision compared to their other eye which underwent microkeratome LASIK (Patel et al. 2007).

Confocal microscopic analysis at 1-week and 1-month post-LASIK demonstrated that the number of reflective particles within the corneal flap interface were similar between eyes who underwent flap creation with IntraLase or mechanical microkeratome (Ramirez et al. 2007). However the nature and origin of these reflective particles remain unclear. Meanwhile, Montes-Mico et al. (2007) demonstrated that LASIK with IntraLase resulted in better contrast sensitivity at high frequencies under photopic and mesopic conditions at 18cpd compared to mechanical microkeratome LASIK ( $p < 0.01$ ). A third comparative study highlighted that the clinical visual results (BSCVA and UCVA) were similar between IntraLase and microkeratome (Moria M2) patients, however IntraLase flaps had more accurate dimensions (i.e. more round flaps and greater



predictability in surface dimensions;  $p=0.016$ ) (Javaloy et al. 2007).

### *Corneal transplantation*

Utilising a new software algorithm (IntraLase Enabled Keratoplasty) for penetrating keratoplasty, Buratto and Bohm (2007) demonstrated that IntraLase achieved nearly clear corneas with a good graft of the donor corneas to the patient's corneas one day post-treatment in all patients ( $n = 7$ , 7 eyes). At three months follow-up, all eyes had a central corneal pachymetry that was within the normal range and a mean thickness of  $\pm 496\mu\text{m}$  (range: 450 to  $521\mu\text{m}$ ). Endothelial cell viability was noteworthy at 3-months, with a mean of  $\pm 2314$  cells/mm (range: 2000 to 2600 cells/mm), however long-term studies would be required in order to demonstrate that this potential advantage of IntraLase is durable in the long-term. Mild corneal haze was observed in one eye and trace haze was observed in another eye at 3-months follow-up. Five patients underwent selective suture removal at 3-months for postoperative astigmatism. Buratto and Bohm (2007) stated that mean refractive astigmatism was 2.9D at 3-months, which is relatively low compared to another (larger) study utilising mechanical microkeratomes where 4D astigmatism was noted at 3-months (Musch et al. 1989). This study provides early clinical evidence that the use of IntraLase for corneal surgery creates a more favourable environment for penetrating keratoplasty. Nevertheless, larger and longer-term studies would be required before the effectiveness of IntraLase in corneal transplantation surgery can be conclusively determined.

One study on corneal endothelial viability after Intralase preparation of posterior lamellar discs from donor eyes for endothelial keratoplasty demonstrated that endothelial cell loss after FS laser lamellar cutting is not dependent on the frequency utilised (15kHz: 7.7%, 30kHz: 4.3%;  $p=0.25$ ) (Cheng et al. 2007). In addition, endothelial cell loss after organ storage was not significant between the IntraLase group and the control group (paired cornea of the same donor) for both 15kHz and 30kHz. However, the investigators were not able to determine the direct effect of the laser on endothelial cell density in the swollen corneas before and after the laser procedure due to procedural limitations (Cheng et al. 2007). In another study, Steinert and colleagues (2007) stated that shaped penetrating keratoplasty with the IntraLase provides superior incision integrity compared to conventional penetrating keratoplasty in donor eyes; indicating that Intralase may provide substantial advantages when utilised for corneal transplantation.

### **Summary of findings**

A considerable amount of new peer-reviewed studies on IntraLase have been published within the last 12 months (seven studies were selected for inclusion). The majority of these studies have focused on the utilisation of IntraLase in LASIK surgery, with a randomised trial (Patel et al. 2007) stating that the lack of clinical differences in vision between mechanical microkeratome LASIK and IntraLase LASIK suggests that both methods are acceptable and appear to be equally safe and effective. Clinical evidence of the effectiveness of IntraLase for corneal transplantation remains limited. However, early indications demonstrate that IntraLase may confer substantial benefits to penetrating keratoplasty (Buratto



and Bohm 2007), while *in vitro* studies have been supportive of its utilisation (Cheng et al. 2007, Steinert et al. 2007).

#### 2007 HealthPACT action

Considering the mounting interest on the use of femtosecond lasers for LASIK and corneal transplantation and its potential for rapid diffusion, it is recommended that IntraLase is monitored for a further 12 months as clinical evidence for its potential application in corneal transplantation are beginning to surface.

- |  |  |
|--|--|
| <input type="checkbox"/> Horizon Scanning Report | <input type="checkbox"/> Full Health Technology Assessment |
| <input checked="" type="checkbox"/> Monitor      | <input type="checkbox"/> Archive                           |

#### Number of studies included

|                                   |   |
|-----------------------------------|---|
| Total number of studies           | 7 |
| Level II intervention evidence    | 1 |
| Level III-2 intervention evidence | 3 |
| Level IV intervention evidence    | 1 |
| Note: <i>In vitro</i> studies     | 2 |

#### References

Buratto L and Bohn E. The use of the femtosecond laser in penetrating keratoplasty. *The American Journal of Ophthalmology* 2007; 143(5): 737-742.

Cheng YYY, Pels E, Cleutjens JPM, van Suylen RJ, Hendrikse F, Nuijts RMMA. Corneal endothelial viability after femtosecond laser preparation of posterior lamellar discs for descemet-stripping endothelial keratoplasty. *Cornea* 26(9): 118-112.

Javaloy J, Vidal MT, Abdelrahman AM, Artola A, Alio JL. Confocal microscopy comparison of IntraLase femtosecond laser and moria M2 microkeratome in LASIK. *Journal of Refractive Surgery* 2007; 23(2): 178-187.

Montes-Mico R, Rodriguez-Galietero, Alio JL, Cervino A. Contrast sensitivity after LASIK flap creation with a femtosecond laser and a mechanical microkeratome. *Journal of Refractive Surgery* 2007; 23(2): 188-192.

Patel SV, Maguire LJ, McLaren JW, Hodge DO, Bourne WM. Femtosecond laser versus mechanical microkeratome for LASIK. *Ophthalmology* 2007; 114(8): 1482-1490.

Ramirez M, Hernandez-Quintela E, Naranjo-Tackman R. A comparative confocal microscopy analysis after LASIK with the IntraLase femtosecond laser vs Hansatome microkeratome. *Journal of Refractive Surgery* 2007; 23(3): 305-307.



Steinert RF, Ignacio TS, Sarayba MA. “Top Hat”-shaped penetrating keratoplasty using the femtosecond laser. *The American Journal of Ophthalmology* 2007; 143(4): 689-691.



## PRIORITISING SUMMARY (2008 UPDATE)

---

|                                  |  |
|----------------------------------|--|
| <b>NAME OF TECHNOLOGY:</b>       | <b>INTRALASE® FEMTOSECOND LASER</b>  |
| <b>PURPOSE AND TARGET GROUP:</b> | <b>THE INTRALASE FEMTOSECOND LASER WAS DEVELOPED TO CREATE CORNEAL FLAPS REQUIRED FOR LASER IN SITU KERATOMILEUSIS (LASIK)</b> |

### **2008 SAFETY AND EFFECTIVENESS ISSUES**

A search of relevant databases, online journals and the Internet was conducted in October 2008, following recommendation in October 2007 that IntraLase femtosecond laser technology for the creation of corneal flaps required for LASIK surgery be monitored for 12 months. Nine new studies (1 randomised controlled trial [RCT], 3 nonrandomised comparative studies, 1 case series and 4 case reports) were identified from this search, two were selected for inclusion.

Of the two included studies, one was a RCT comparing corneal wavefront aberration changes following LASIK performed with IntraLase femtosecond laser (13 patients [23 eyes]) and Hansatome microkeratome (15 patients [24 eyes]) (Buzzonetti et al 2008). The other was a nonrandomized comparative study looking at the incidence of diffuse lamellar keratitis (DLK) in consecutive patients following LASIK with IntraLase 15 kHz femtosecond laser (1000 eyes) and Moria M2 microkeratome (1000 eyes) (Gil-Cazorla et al 2008).

Of the nonrandomised comparative studies not included in this update, one compared LASIK using IntraLase femtosecond laser for flap creation with mechanical microkeratome for flap creation in regards to residual refractive error correction after refractive lens exchange (RLE) with multifocal intraocular lens (IOL) implantation (Alfonso et al 2008) and the other in regards to changes in the corneal curvature from the predicted surgical radius (Ortiz et al 2008). Alfonso et al (2008) concluded femtosecond laser after RLE with multifocal IOL implantation is an effective procedure for correcting residual ametropia and resulted in better distance visual outcomes and Ortiz et al (2008) found the refractive change in corneal curvature was lower after femtosecond laser LASIK than after LASIK performed using a mechanical microkeratome. These studies also recognised the need for further studies to support their findings, including those to assess patient satisfaction and those with longer follow-up duration to confirm the safety of the procedure (Alfonso et al 2008).

### **Safety**

There was no safety data reported in the study by Buzzonetti et al (2008).

Gil-Cazorla et al (2008) reported uneventful surgery in all 2000 eyes receiving LASIK, both in the IntraLase femtosecond laser group and the microkeratome group.

### **Effectiveness**

Buzzonetti et al (2008) reported the postoperative mean spherical equivalent defect at  $-0.3 \pm 0.8$  diopters (D) in the IntraLase group and  $-0.18 \pm 0.5$  D in the microkeratome group. The postoperative mean uncorrected visual acuity was  $8.6 \pm 2.3$  in the IntraLase group and  $9.3 \pm 1.6$  in the microkeratome group, whereas the mean best-corrected visual acuity was  $9.1 \pm 1.1$  in the IntraLase group and  $9.5 \pm 0.9$  in the microkeratome group. The postoperative mean



decentration of ablation was not significantly different between the two groups, at  $<0.5\text{mm}$ . The comalike aberration appeared to be positively correlated with decentration of ablation in both groups ( $p < 0.05$ ). The 3.0 mm pupil significantly changed the amount of comalike aberration in the microkeratome group ( $p < 0.05$ ), in contrast, with the 5.0 mm pupil the amount of total high-order corneal wavefront aberrations significantly changed in both groups ( $p < 0.05$ ).

Increasing factor (IF) was calculated. This was the ratio between the postoperative and preoperative mean value of the optical aberrations. At three months follow-up, with the 3.0 mm pupil in the IntraLase group, the IF for total aberration was 1.3, for spherical-like aberration 2.0 and for comalike aberration 1.7. In the microkeratome group, the IF for total aberration was 1.1, for spherical-like aberration 1.8 and for comalike aberration 1.4. At one year postoperative in the IntraLase group, the IF for total aberration was 1.3, for spherical-like aberration 1.2 and for comalike aberration 1.5. In the microkeratome group at one year follow-up, the IF for total aberration was 2.0, for spherical-like aberration 1.2 and for comalike aberration 2.1 (Buzzonetti et al 2008).

At three months follow-up with the 5.0 mm pupil in the IntraLase group, the IF for total aberration was 1.8, for spherical-like aberration 2.6 and for comalike aberration 2.8. In the microkeratome group, the IF for total aberration was 1.7, for spherical-like aberration 3.18 and for comalike aberration 1.7. At one year postoperative in the IntraLase group, the IF for total aberration was 1.7, for spherical-like aberration 2.4 and for comalike aberration 1.7. In the microkeratome group at one year follow-up, the IF for total aberration was 3.3, for spherical-like aberration 2.9 and for comalike aberration 2.7 (Buzzonetti et al 2008).

The study by Buzzonetti et al (2008) concluded that wavefront corneal aberrations changed significantly at one-year following LASIK surgery with both IntraLase femtosecond laser and microkeratome. Both procedures induced a similar degree of spherical-like aberration in the anterior corneal surface. Comalike aberrations were increased following microkeratome LASIK compared with IntraLase LASIK. Therefore because induced aberrations may interfere with custom ablation success, IntraLase may provide more effective wavefront-guided corrections (Buzzonetti et al 2008).

In the study by Gil-Cazorla et al (2008), stage 2 DLK occurred in 0.1% (1/1000) of patients in both the IntraLase femtosecond laser group and the microkeratome group. Stage 3 DLK occurred in 0.4% (4/1000) of patients from the IntraLase femtosecond laser group only. There was no need for flap lifting and interface irrigation in either of the groups and all cases of DLK resolved without serious corneal scarring. Overall, the incidence of postoperative DLK was significantly higher in the IntraLase group (0.5%) compared with the microkeratome group (0.1%) ( $P=0.03$ ).

## **2008 RECOMMENDATION**

The available literature regarding IntraLase femtosecond laser use for the creation of corneal flaps in LASIK surgery remains inconclusive. The included studies were unable to prove superiority of IntraLase over conventional microkeratome techniques in regards to aberration changes and DLK occurrence, and encouraged further investigation regarding the procedure, with particular emphasis on new versions of IntraLase (30 and 60 kHz software), in order to reliably compare the efficacy of the procedures. Based on the evidence to date, this technology will be archived due to inconclusive evidence.



## 2008 NUMBER OF STUDIES INCLUDED

|                         |   |
|-------------------------|---|
| Total number of studies | 2 |
| Level II evidence       | 1 |
| Level III evidence      | 1 |

## 2008 REFERENCES

- Alfonso JF, Fernandez-Vega L, Montes-Mico R, Valcarcel B. Femtosecond laser for residual refractive error correction after refractive lens exchange with multifocal intraocular lens implantation. *Am J Ophthalmol* 2008; **146**(2): 244-250.
- Buzzonetti L, Petrocelli G, Valente P, Tamburrelli C, Mosca L, Laborante A, Balestrazzi E. Comparison of corneal aberration changes after laser in situ keratomileusis performed with mechanical microkeratome and IntraLase femtosecond laser: 1-year follow-up. *Cornea* 2008; **27**(2): 174-179.
- Choi SK, Kim JH, Lee D, Oh SH, Lee JH, Ahn MS. Creation of an extremely thin flap using IntraLase femtosecond laser. *J Cataract Refract Surg* 2008; **34**(5): 864-867.
- Gil-Cazorla R, Teus MA, Benito-Llopis L, Fuentes I. Incidence of diffuse lamellar keratitis after laser in situ keratomileusis associated with the IntraLase 15 kHz femtosecond laser and Moria M2 microkeratome. *J Cataract Refract Surg* 2008; **34**(1): 28-31.
- Kaiserman I, Maresky HS, Bahar I, Rootman DS. Incidence, possible risk factors, and potential effects of an opaque bubble layer created by a femtosecond laser. *J Cataract Refract Surg* 2008; **34**(3): 417-423.
- Krueger RR and Dupps WJ, Jr. Biomechanical effects of femtosecond and microkeratome-based flap creation: prospective contralateral examination of two patients. *J Refract Surg* 2007; **23** (8): 800-807.
- Ortiz D, Alio JL, Pinero D. Measurement of corneal curvature change after mechanical laser in situ keratomileusis flap creation and femtosecond laser flap creation. *J Cataract Refract Surg* 2008; **34**(2): 238-242.
- Seider MI, Ide T, Kymionis GD, Culbertson WW, O'Brien TP, Yoo SH. Epithelial breakthrough during IntraLase flap creation for laser in situ keratomileusis. *J Cataract Refract Surg* 2008; **34**(5): 859-863.
- Srinivasan S and Rootman DS. Anterior chamber gas bubble formation during femtosecond laser flap creation for LASIK. *J Refract Surg* 2007; **23**(8): 828-830.